



```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                      Location  : P2-B-01
Injection Date  : 7/21/2022 6:05:02 AM        Inj       :    2
                                           Inj Volume: 2.000 µl
```

```
Acq. Method    : C:\Users\Public\Documents\ChemStation\1\Data\07212201\07212201 2022-07-21
05-45-17\PFAA.M
```

```
Last changed   : 1/26/2021 10:30:23 AM by SYSTEM
```

```
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
```

```
Last changed   : 3/23/2022 11:04:54 AM by SYSTEM
```

```
Method Info    : OPERATOR: Organic QC lab
```

```
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity
II LC/MS system, Agilent 6125 quadrupole MS detector.
```

```
Additional Info : Peak(s) manually integrated
```

Current Chromatogram(s)

