



=====

Acq. Operator : SYSTEM	Seq. Line : 19
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P2-A-09
Injection Date : 2/14/2023 2:01:51 PM	Inj : 1
	Inj Volume : 2.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.200 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\02142301\021423101 2023-02-14 08-21-15\PFAA.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Data\01302301\013023101 2023-01-30 06-52-36\Purity_195-315.M (Sequence Method)

Last changed : 2/14/2023 9:29:08 AM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

