

File :Q:\GCMS-06 Minimal\DATA\020421\06- -DC.D  
Operator : Organic QC Lab  
Acquired : 04 Feb 2021 11:18 using AcqMethod CHICK\_2019\_13MIN.M  
Instrument : GCMS 6  
Sample Name: PFRS-025S  
Misc Info : RT: tris(1-Chloro-2-propyl) phosphate (TCPP)L in Toluene  
Vial Number: 46



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

