

File :D:\MassHunter\GCMS\1\DATA\122221\21- ALR.D
Operator : Organic QC Lab
Acquired : 22 Dec 2021 14:18 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: PHTH-0205
Misc Info : Di-n-heptyl phthalate in MeOH
Vial Number: 49



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

