

File :Q:\GCMS-06 Minimal\DATA\053119\BSB\14-.D
Operator : D.C
Acquired : 31 May 2019 22:24 using AcqMethod CHICK_2017_S2_2uL.M
Instrument : GCMS 6
Sample Name: PHTH-0255-H
Misc Info : E0: Diisotridecyl phthalate in Hexane
Vial Number: 68



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 2:1, 2uL inj.; GC/MS; INJ 270c

