

File :Q:\GCMS-08\DATA\20190225\04- D
Operator : JP/KM
Acquired : 25 Feb 2019 17:47 using AcqMethod KERLA-2017-100uL draw speed.M
Instrument : GCMS 8
Sample Name: PHTH-D4-005S
Misc Info : Diethyl phthalate-3,4,5,6-d4 in MeOH
Vial Number: 56



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

