

File :D:\MassHunter\GCMS\1\data\20220712\11 - SAK.D
Operator : Organic QC Lab
Acquired : 12 Jul 2022 11:35 using AcqMethod Hy_Pest_24cm_S10.M
Instrument : GCMS 8
Sample Name: PHTH-D4-009S
Misc Info : Di-n-pentyl Phthalate-3,4,5,6-d4
Vial Number: 123



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

