

File : Q:\GCMS-08 Minimal\DATA\20210519\14- [JP].D
Operator : Organic QC Lab
Acquired : 19 May 21 03:45 pm using AcqMethod Hy_Pest_15cm_S50.M
Instrument : GCMS 8
Sample Name: PLAS-AX-018S
Misc Info : Isonox 132; in hexane
Vial Number: 59



Column: DB-5, 30m, 0.25ID, 0.25um
Oven: 80c, 2m, 30c/m to 320, final 2min
GC para: Pulse split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj, GCMS, Inj 280c

