

File :Q:\GCMS-08 Minimal\DATA\20240119\50- \_JP~.D  
Operator : Organic QC Lab  
Acquired : 20 Jan 24 03:21 am using AcqMethod Hy\_BDE\_24cm\_S20.M  
Instrument : GCMS 8  
Sample Name: PLAS-AX-021S  
Misc Info : Ethanox 330; in hex  
Vial Number: 26



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

