

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20221031
... \47- KM.D
Operator : Organic QC Lab
Instrument : GCMS 8
Acquired : 01 Nov 2022 03:37 using AcqMethod Hy_Pest_15cm_S10.M
Sample Name: PLAS-AX-022S
Misc Info : Naugard 445 in Hexane



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

