

File :D:\MassHunter\GCMS\1\data\20210901\14 JP {.D
Operator : Organic QC Lab
Acquired : 01 Sep 2021 13:56 using AcqMethod Hy_BDE209_S5.M
Instrument : GCMS 8
Sample Name: PLAS-AX-045S
Misc Info : Irganox 259; in hexane
Vial Number: 91



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

