

File :Q:\GCMS-08 Minimal\DATA\20210901\16- JP=.D  
Operator : Organic QC Lab  
Acquired : 01 Sep 2021 16:19 using AcqMethod Hy\_BDE209\_S5.M  
Instrument : GCMS 8  
Sample Name: PLAS-AX-059S  
Misc Info : Lowinox CPL  
Vial Number: 53



Column: DB-5, 30m, 0.25ID, 0.25um  
Oven: 80c, 2m, 30c/m to 320, final 2min  
GC para: Pulse split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj, GCMS, Inj 280c

