

File : Q:\GCMS-08 Minimal\DATA\20230628\12-[ JP ].D  
Operator : Organic QC Lab  
Acquired : 28 Jun 2023 16:25 using AcqMethod Hy\_Pest\_24cm\_S10.M  
Instrument : GCMS 8  
Sample Name: PLAS-CA-004S  
Misc Info : Silquest A-187; in hexane  
Vial Number: 7



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; injector 270°C

