

File : Q:\GCMS-07\DATA\03061901\16- D
Operator : K.D
Acquired : 06 Mar 2019 15:20 using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name: PLAS-CPSC-R1
Misc Info : in Cyclohexane
Vial Number: 138

Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

