

File :D:\MassHunter\GCMS\1\data\06202022\13- ALR.D  
Operator : ALR  
Acquired : 20 Jun 2022 12:15 using AcqMethod Hy SV DB-5MS UI20M 0.18 50 split FAST .M  
Instrument : GCMS 9  
Sample Name: PLAS-PL-091S (217121187)  
Misc Info : RT: Di-n-butyl maleate in Hexane  
Vial Number: 71



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

