

File : Q:\GCMS-08 Minimal\DATA\20230224\25- _JP{.D
Operator : Organic QC Lab
Acquired : 24 Feb 23 11:33 pm using AcqMethod Hy_Pest_24cm_S50.M
Instrument : GCMS 8
Sample Name: PLAS-PL-104S
Misc Info : Tris(2-butoxyethyl) phosphate; in hexane
Vial Number: 11



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

