

File :Q:\GCMS-08 Minimal\DATA\20230314\04- }JP=-.D
Operator : Organic QC Lab
Acquired : 14 Mar 23 04:06 pm using AcqMethod Hy_Pest_15cm_S10.M
Instrument : GCMS 8
Sample Name: PLAS-RT-001S
Misc Info : Retarder AK; in hex:acet [60:40]
Vial Number: 4



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

