

File :D:\MassHunter\GCMS\1\DATA\021522\02- -KD.D
Operator : Organic QC Lab
Acquired : 15 Feb 2022 06:37 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: PLAS-UV-015S-CN
Misc Info :
Vial Number: 148



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

