

File :D:\MassHunter\GCMS\1\DATA\091522\04 - SAK.D  
 Operator : Organic QC Lab  
 Acquired : 15 Sep 2022 09:16 using AcqMethod CHICK\_2019\_VOC-SV.M  
 Instrument : GCMS 6  
 Sample Name: PS-CP-06A-1ML  
 Misc Info : Alkanes C8-C40 in Hex  
 Vial Number: 141

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
 Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
 GC Parameters: Pulsed split 40psi 0.5m in, 10psi constant flow  
 Split 2:1, 2µl inj.; GC/MS; Injector 270°C

