

File :Q:\GCMS-07 Minimal\DATA\08142001\08- KM.D
Operator : Organic QC department
Acquired : 14 Aug 2020 10:44 using AcqMethod KERLA-2Split-2uL.M
Instrument : 5977B MSD
Sample Name: RAC-06-EA-0.1X-10ML
Misc Info : 4-Chloro-o-toluidine in Ethyl acet
Vial Number: 48



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

