

File :Q:\GCMS-08 Minimal\DATA\20221219\42- 222121327 =JP=.D
Operator : Organic QC Lab
Acquired : 19 Dec 22 08:59 pm using AcqMethod Hy_HC08-HC40_S100.M
Instrument : GCMS 8
Sample Name: RRO-AK-103AA
Misc Info : in dcm
Vial Number: 2



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

