

File :Q:\GCMS-04\DATA\021519\10-.D
Operator : RFXM
Acquired : 15 Feb 2019 18:57 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: S-406A-10X
Misc Info : RT
Vial Number: 45



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

