


File : L:\QC Data\GCMS-06 Minimal\DATA\062323\4- sak.D  
Operator : Organic QC Lab  
Acquired : 24 Jun 23 02:51 am using AcqMethod CHICK\_2019\_VOC-SV.M  
Instrument : GCMS 6  
Sample Name: SRD-0.1X-100ML  
Misc Info :  
Vial Number: 42



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 35°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C



RT Component  
5.451 Sulfur (7704-34-9)\*  
\* as Di-n-butyl sulfide (544-40-1)

