

File : L:\QC Data\GCMS-06 Minimal\DATA\062323\4- sak.D
Operator : Organic QC Lab
Acquired : 24 Jun 23 02:51 am using AcqMethod CHICK_2019_VOC-SV.M
Instrument : GCMS 6
Sample Name : SRD-3X-100ML
Misc Info :
Vial Number : 42



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 35°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

