

File : L:\QC Data\GCMS-06 Minimal\DATA\062323\53 - sak.D
Operator : Organic QC Lab
Acquired : 24 Jun 2023 08:19 using AcqMethod CHICK_2019_VOC-SV_S100.M
Instrument : GCMS 6
Sample Name : SWMO-200X-100ML
Misc Info : diluted
Vial Number : 46



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 100:1, 1µl inj.; GC/MS; Injector temp 270°C

