

File :D:\MassHunter\GCMS\1\DATA\11012201\43 - SAK.D  
Operator : Organic QC department  
Acquired : 02 Nov 2022 06:04 using AcqMethod KERLA-2017\_Short\_15Min\_S100.M  
Instrument : 5977B MSD  
Sample Name: UFA-030S  
Misc Info : dil 10x  
Vial Number: 139



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

