

File :Q:\GCMS-06 Minimal\DATA\020921\19- KM.D  
Operator : Organic QC Lab  
Acquired : 09 Feb 2021 19:13 using AcqMethod CHICK\_2019\_13MIN.M  
Instrument : GCMS 6  
Sample Name: UOP-M-660-0.1X  
Misc Info : UOP Method 660 Standard in Water  
Vial Number: 108



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

